



Part of Ant Group Ltd

## Latex Medical Examination Gloves

### Features:

- Material specification: Natural rubber latex
- Manufacturing process types: Online chlorination on donning surface and polymer coating on working surface
- Colour: Off white to yellow
- Texture and grip: Full palm-textured surface finished, good grip
- Former type: E5 full palm-textured former



Single Use



### Latex Caution Statement:

Safe use of this glove by or on latex sensitized individuals has not been established.  
This product contains natural rubber latex which may cause allergic reactions.



### Technical Specification:

#### Freedom from holes BS EN 455-1 : 2000

Water Tight Test: General Inspection Level 1, AQL 1.5

#### Biological Evaluation BS EN 455-3 : 2000

Protein Test:  $\leq 50\mu\text{g/g}$  using the modified Lowry method

#### Residual Powder ASTM D6124-01

Powder Content: Samples should contain no more than 1.0mg of any filterable residue per glove

### Quality Standards (ISO BS EN) Expectation

Manufacturing QMS:	BS EN ISO 9001 : 2000
Glove:	BS EN 455 Parts 1, 2 & 3
Sampling:	ISO 2859-1
Medical Device Directive Requirement:	93/42/EEC Class 1
PPE Directive Requirement:	N/A
Technical File Requirements & Availability:	Available for inspection by Brosch direct or their notified body
Shelf Life:	3 years

Product Code:	Size:	Width (mm):	Length (mm):	Tensile Strength:	Double Wall Thickness:
509-0015	Small	$80 \pm 10$	220min	Before aging: min. 9N After aging: min. 6N	Finger: 0.15mm Palm: 0.13mm
509-0016	Medium	$95 \pm 10$	230min		
509-0017	Large	$110 \pm 10$	230min		
509-0018	Extra Large	$\geq 110$	230min		

Product Code:	Description:	Size:	Additional Notes:
509-0015	Latex Medical Examination Gloves	Small	Box of 100
509-0016	Latex Medical Examination Gloves	Medium	Box of 100
509-0017	Latex Medical Examination Gloves	Large	Box of 100
509-0018	Latex Medical Examination Gloves	Extra Large	Box of 100